



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST (h) PAI-1 Antigen

Product Code : RK012A

Lot No. : FA2078

Expiration date : 2024-05-13

Components	Quantity	Lot	Expiration Date
Anti-(h)-PAI-1 pre-coated plate	12x8 wells	FA2093	2025-04-21
(h) PAI-1 standard	3 vials	FA2111	2025-04-28
(h) PAI-1 Control I	1 vial	FA2112	2025-04-28
(h) PAI-1 Control II	1 vial	FA2110	2025-04-28
Anti-(h)-PAI-1 HRP immunoconjugate	3 vials	FA2216	2025-05-15
F-Sample diluent	2x50 ml	F2001620	2024-05-13
Wash solution	1x50 ml	FA2213	2025-05-12
Conjugate diluent	1x25 ml	FA1121	2024-11-18
TMB	1x25 ml	210302D01	2025-03-30
Sulfuric Acid 0,45M	1x6 ml	FA1281	2024-12-24

Analytical Data	Specification	Result
Reactivity	A450 for 8 ng/ml \geq 1.40	1.56
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.003
Detection threshold	\leq 0.5 ng/ml	< 0.006

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	1.978	1.225	0.696	0.335	0.158

MV = Measured value

	Acceptance Range	MV
CI	[37.7 - 50.9]	47.4
CII	[13.3 - 19.9]	17.3

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 09/12/2021

Quality Control Manager :

S.LECOURT

