



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST Fibrinogen

Product Code : RK024A

Lot No. : FA2177

Expiration date : 2024-05-10

Components	Quantity	Lot	Expiration Date
pre-coated plate	12x8 wells	FA2209	2024-05-17
Fibrinogen Standard	3 vials	FA2231	2024-05-24
Fibrinogen Control I	1 vial	FA2229	2024-05-24
Fibrinogen Control II	1 vial	FA2230	2024-05-24
Fibrinogen HRP immunoconjugate	3 vials	FB0086	2024-07-01
B2F Sample diluent	2x50 ml	FA2169	2024-05-10
Wash solution	1x50 ml	FA1511	2025-02-10
B2F Conjugate diluent	1x25 ml	FA2170	2024-05-10
TMB	1x25 ml	210302D01	2025-03-30
Sulfuric Acid 0,45M	1x6 ml	FA0937	2024-10-14

Analytical Data	Specification	Result
Reactivity	A450 for 40 ng/ml \geq 1.45	1.75
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.150	0.030
Detection threshold	\leq 1 ng/ml	< 0.5 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-01-27

Quality Control Manager : S.LECOURT