

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST Activatable TAFI

Product Code : RK037A

Lot No. : FA1744

Expiration date : 2024-05-31

| Components | Quantity | Lot | Expiration Date |
|----------------------------------------|------------|-----------|-----------------|
| Anti-(h) Act TAFI pre-coated plate | 12x8 wells | FB0032 | 2024-06-23 |
| TAFI standard | 3 vials | FA2310 | 2024-05-31 |
| Control I | 1 vial | FA2306 | 2024-05-31 |
| Control II | 1 vial | FA2307 | 2024-05-31 |
| Anti-(h)- Act TAFI HRP immunoconjugate | 3 vials | FB0075 | 2024-07-01 |
| Fibrinolysis Sample diluent | 2x50 ml | FA1742 | 2025-04-21 |
| Wash solution | 1x50 ml | FA0257 | 2024-07-03 |
| Conjugate diluent | 1x25 ml | FA1119 | 2024-11-18 |
| TMB | 1x25 ml | 210302D01 | 2025-03-30 |
| Sulfuric Acid 0,45M | 1x6 ml | F2001723 | 2024-06-19 |

| Analytical Data | Specification | Result |
|--------------------------------|---------------------------------------------------|--------|
| Reactivity | A450 for 100% \geq 1.40 | 1.64 |
| Calibration Curve and Accuracy | Controls Measured Values within acceptance ranges | PASS |
| Blank | < 0.100 | 0.007 |
| Detection threshold | \leq 10 % | < 2% |

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-02-02

Quality Control Manager : S.LECOURT