



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CALIBRATOR

Product Code : 222701

Lot No. : FA229519

Expiration date : 2024-06-04

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FA2295BA19	2024-06-04
CAL 2	4 vials	FA2295BB19	2024-06-04
CAL 3	4 vials	FA2295BC19	2024-06-04

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/ml CAL 2 : 200 -300 ng/ml CAL 3 : 400 -600 ng/ml	0 ng/mL 263 ng/mL 500 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-01-17

Quality Control Manager : S.LECOURT

