

# Purified bovine Factor Xa

- REF BE101B  Ez Vial of 0.5 mg/Flacon de 0.5 mg
- REF BE101C  Ez Vial of 15 µg/flacon de 15 µg
- REF BE101D  Ez Vial of 30 µg /Flacon de 30 µg
- REF BE101L  Ez Set of 10x30 µg/Coffret de 10x30 µg
- REF BE101O  Ez Vial of 50 µg /Flacon de 50 µg
- REF BE101K  Ez Set of 6x50 µg/Coffret de 6x50 µg

STORE AT  
2-8°C

CONSERVER A  
2-8°C

REF BE101B

REF BE101C

REF BE101D/L

REF BE101O/K

LOT F2000851

EXP 2024-06-01

## ANALYSIS CERTIFICATE / CERTIFICAT D'ANALYSE

ANALYTICAL DATA/RESULTATS D'ANALYSE			SPECIFICATIONS	
1.	<u>Enzymatic activity on Factor Xa Chromogenic Substrate CS-11(65) / Activité enzymatique sur substrat chromogène CS-11(65) du FXa</u>		≥ 50 U(*) ≥ 150 Nkats	
		Nkats		U (*)
	Factor Xa Potency	211		61
2.	<u>Batch homogeneity / Homogénéité de lot :</u>		CV ≤ 3%	
	N = 15	CV : 1.17 %		

(\*)Standardization: Purified bovine factor Xa is standardized against WHO reference reagent 15/102 from NIBSC

### CONCLUSIONS

DATE: 2020-07-02

Passed/Conforme

Refused/Non-conforme

Quality Control Manager  
Responsable contrôle qualité

S. LECOURT