



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : PT-Phen™ LRT
Product Code : CK586K-CK586L
Lot No. : FB1117AJ
Expiration date : 2024-06-29

Components	Quantity	Lot	Expiration Date
R1 : PT-Phen™ LRT	6 vials (CK586K) 12 vials (CK586L)	FB1117doAJ	2024-06-29

Analytical Data	Specification	Result
PT on Normal Plasma	11 ± 1 sec	11.3 sec (on CS)
ISI (WHO Method)	1.0 ± 0.1	0.98 sec (on CS)

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-09-12

Quality Control Manager :

Manufacturing date : 2022-07-13