



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Item Description : LIAPHEN AT

Item Code : 120002

Lot No. : FB00628Z

Expiration date : 2024-07-12

Components	Quantity	Lot	Expiration Date
R1 : Latex reagent	2 vials	FB0062IW8Z	2024-07-12
R2 : Reaction Buffer	2 vials	FB0062IX8Z	2024-07-12

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	A620 (0-150%) ≥ 0.40 $R^2 \geq 0.98$ Controls measured within acceptance ranges	0.67 PASS PASS
Detection threshold	$\leq 10\%$	$< 7\%$
Hook effect	$> 200\%$	$> 300\%$

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-02-08

Quality Control Manager : S.LECOURT