



A Company of Sysmex Group

HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Item Description: LIAPHEN Fibrinogen	
Item Code: 120102	
Lot No.: FB00661F	
Expiration date: 2024-07-26	

Components	Quantity	Lot	Expiration Date
R1 : Latex reagent	4 vials	FB0066JA1F	2024-07-26

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	Delta A620 (0-20µg/ml) ≥ 0.40 Controls measured within acceptance ranges R ² ≥ 0.98	0.740 PASS PASS
Limit of Quantification	≤ 1 µg/ml (in the test dilution (1:200))	< 1 µg/mL
Hook effect	> 90 µg/ml (in the test dilution (1:200))	> 90 µg/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-04-22

Quality Control Manager: S.LECOURT