



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description:** BIOPHEN™ Dabigatran Plasma Calibrator  
**Product Code:** 222801  
**Lot No.:** FB01461A  
**Expiration date:** 2024-07-15

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FB0146AS1A	2024-07-15
CAL 2	4 vials	FB0146AT1A	2024-07-15
CAL 3	4 vials	FB0146AU1A	2024-07-15

Analytical Data	Specification	Result
Concentration	CAL 1: $\leq 100$ ng/ml CAL 2: 150 -350 ng/ml CAL 3: 400 -600 ng/ml	45 274 516
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2022-02-11

Quality Control Manager: S.LECOURT