



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS
Product Description : ZYMUTEST HIA IgG,A,M
Product Code : RK040D
Lot No. : FB0088
Expiration date : 2024-07-02

Components	Quantity	Lot	Expiration Date
Heparin pre-coated plate	12x8 wells	FB0123	2024-07-05
HIA Ig G,A,M Positive Control	3 vials	FB0316	2024-08-05
Negative Control	3 vials	FA2223	2025-05-13
Anti (h) IgG - Anti IgA - Anti IgM HRP immunoconjugate	3 vials	FB0278	2025-07-24
HIA Sample diluent	2x50 ml	FB0124	2024-07-02
Wash solution	1x50 ml	FA2213	2025-05-12
Cell Lysate	3 vials	FB0229	2024-07-19
Conjugate diluent	1x25 ml	FA1121	2024-11-18
TMB	1x25 ml	210302D01	2025-03-30
Sulfuric Acid 0,45M	1x6 ml	FB0237	2025-07-16

Analytical Data	Specification	Result
Reactivity	A450 for Positive Control ≥ 1.00 A450 for Negative Control ≤ 0.30	1.90 0.09
Pathological Plasmas	A450 ≥ 0.500	PASS
Blank	< 0.100	0.022
Normal Plasma	M + 2SD ≤ 0.50	0.266

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-03-07

Quality Control Manager :

S.LECOURT
