



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ARGATROBAN CONTROL PLASMA
Product Code : SC035K
Lot No. : FB006417
Expiration date : 2024-07-02

Components	Quantity	Lot	Expiration Date
C1	6 vials	FB0064CE17	2024-07-02
C2	6 vials	FB0064CF17	2024-07-02

Analytical Data	Specification	Result
C1	0.55 - 0.80 µg/mL	0.71
C2	1.10 - 1.40 µg/mL	1.38

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-02-02

Quality Control Manager : S.LECOURT

