



 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Item Description : LIAPHEN™ Protein C	
Item Code : 120003	
Lot No. : FC071814	
Expiration date : 2024-08-13	

Components	Quantity	Lot	Expiration Date
R1 : Reaction Buffer	4 vials	FC0718en14	2024-08-13
R2 : Latex reagent	4 vials	FC0718eo14	2024-08-13

Analytical Data	Specification	Result
Calibration curve : reactivity, slope and accuracy	OD/min (3/2) ≥ 0.30 Delta OD/min (1/1-1/2) ≥ 0.1 Controls measured within acceptance ranges	0.393 0.139 PASS
Hook effect	OD/min (300 et 600 %) ≥ OD/min (point 1/1 of calibration curve)	600%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-04-20

Quality Control Manager : S.LECOURT

P/O 