



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN RIVAROXABAN CALIBRATOR
Product Code: 222701
Lot No.: FB03721A
Expiration date: 2024-08-12

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FB0372BA1A	2024-08-12
CAL 2	4 vials	FB0372BB1A	2024-08-12
CAL 3	4 vials	FB0372BC1A	2024-08-12

Analytical Data	Specification	Result
Concentration	CAL 1: ≤ 50 ng/mL CAL 2: 200 -300 ng/mL CAL 3: 400 -600 ng/mL	0 279 453
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-03-10

Quality Control Manager: S. LECOURT

