



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN UFH CONTROL C2
Product Code : 223901
Lot No. : FB02981A
Expiration date : 2024-08-03

Components	Quantity	Lot	Expiration Date
C2	12 vials	FB0298BY1A	2024-08-03

Analytical Data	Specification	Result
Concentration	C2 : 0.40 -0.60 IU/mL	0.53

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-03-08

Quality Control Manager :

S.LECOURT