



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN APIXABAN CALIBRATOR LOW
Product Code: 226101
Lot No.: FB03781A
Expiration date: 2024-08-10

Components	Quantity	Lot	Expiration Date
CAL I	4 vials	FB0378DT1A	2024-08-10
CAL II	4 vials	FB0378DU1A	2024-08-10
CAL III	4 vials	FB0378DV1A	2024-08-10

Analytical Data	Specification	Result
Concentration	CAL I : < 20 ng/mL CAL II : 40 -60 ng/mL CAL III : 80 -120 ng/mL	0 ng/mL 54 ng/mL 104 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-04-21

Quality Control Manager: S. LECOURT

