



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: EPINEPHRINE
Product Code: AG002K
Lot No.: FB0312
Expiration date: 2024-08-05

Components	Quantity	Lot	Expiration Date
EPINEPHRINE	3 vials	FB0312	2024-08-05

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 5 µM on CS	≥ 70%	91%
Reactivity (% AgregMax) at 5 µM on Agregometer	≥ 70%	105%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-03-29

Quality Control Manager: S. LECOURT

