



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** IMIDAZOLE BUFFER  
**Product Code :** AR021K/L  
**Lot No. :** FB03639D  
**Expiration date :** 2024-08-09

Components	Quantity	Lot	Expiration Date
IMIDAZOLE BUFFER	3 vials *25mL (AR021K) 12 vials *25mL (AR021L)	FB0363AR9D	2024-08-09

Analytical Data	Specification	Result
Hemoclot PC clotting assay	Delta CT (100%) with control sample $\leq$ 5%	< 1 %
FVIII:C clotting assay	Delta CT (100%) with control sample $\leq$ 5%	< 1%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2022-03-08

Quality Control Manager : S.LECOURT

