



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Item Description: LIAPHEN AT

Item Code: 120002

Lot No.: FB04829N

Expiration date: 2024-09-13

| Components | Quantity | Lot | Expiration Date |
|----------------------|----------|------------|-----------------|
| R1 : Latex reagent | 2 vials | FB0482IW9N | 2024-09-13 |
| R2 : Reaction Buffer | 2 vials | FB0482IX9N | 2024-09-13 |

| Analytical Data | Specification | Result |
|---|--|------------------------|
| Calibration curve : reactivity, slope, linearity and accuracy | A620 (0-150%) ≥ 0.40 $R^2 \geq 0.98$ Controls measured within acceptance ranges | 0.794 1.000 PASS |
| Detection threshold | $\leq 10\%$ | < 6% |
| Hook effect | >200% | > 300% |

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-04-19

Quality Control Manager: S.LECOURT

P.O.