

Certificate of Analysis

Product: 5-PLASMA Normal Plasma 2

Reference Number: 5D-40401A

Lot Number: 244609

Expiration Date: 2024-09

Storage: 2-8°C


Presentation: 10 vials of 2 mL

Specifications:

Analytical Data	Specifications	Result
INR	< 1.3	1.20
aPT ratio	< 1.20	1.10
Fibrinogen	≥ 2 g/L	2.25 g/L
FVIII:C	> 50%	76%
LA DVV test / DVV confirm ratio	≤ 1.20	1.06

For Research Use Only. Not for use in diagnostic procedures. For *in vitro* use only.

Signature:



Urs-Peter Geiger
Date: 2022-11-03