

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN PLASMA CALIBRATOR

Product Code: 222101

Lot No.: FB06641A

Expiration date: 2024-10-07

Components	Quantity	Lot	Expiration Date
Plasma Calibrator	12 vials	FB0664DY1A	2024-10-07

Analytical Data	Specification	Result
Fibrinogen (Activity)	> 2 g/L	2.83 g/L
Fibrinogen (Antigen)	> 2 g/L	2.85 g/L
Alpha 2 Anti plasmin	> 85 %	96%
Anti-thrombin (Activity)	> 85 %	92%
Anti-thrombin (Antigen)	> 85 %	91%
Factor II	> 80 %	89%
Factor V	> 80 %	91%
Factor VII	> 80 %	94%
Factor X	> 80 %	81%
Factor VII+ X	> 85 %	89%
Factor VIII	> 60 %	104%
Factor IX	> 60 %	86%
Factor XI	> 70 %	84%
Factor XII	> 70 %	86%
Factor XIII	> 70 %	99%
Free Protein S (Antigen)	70 - 140 %	138%
Plasminogen	> 80 %	98%
Protein C Chromogenic	> 85 %	92%
Protein C Clotting	> 70 %	97%
Protein C (Antigen)	> 75 %	82%
Protein S (Activity)	> 50 %	119%
von Willebrand Factor (Antigen)	> 60 %	110%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-07-19

Quality Control Manager: S. LECOURT

