



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN ABNORMAL CONTROL PLASMA

Product Code: 223301

Lot No.: FB066819

Expiration date: 2024-10-11

| Components | Quantity | Lot | Expiration Date |
|------------------|----------|------------|-----------------|
| Abnormal Control | 12 vials | FB0668EA19 | 2024-10-11 |

| Analytical Data | Specification | Result |
|---------------------------------|---------------|----------|
| INR | > 1.3 | > 1.3 |
| aPTT LR Ratio | > 1.50 | > 1.5 |
| Fibrinogen (Activity) | 0.8-1.2 g/L | 1.08 g/L |
| Fibrinogen (Antigen) | 0.7-1.3 g/L | 1.13 g/L |
| LA ratio | ≤ 1.2 | < 1.2 |
| Alpha 2 Anti plasmin | 20 - 50 % | 36% |
| Antithrombin (Activity) | 20 - 50 % | 35% |
| Antithrombin (Antigen) | 20 - 50 % | 36% |
| Factor II | 20 - 50 % | 35% |
| Factor V | 20 - 50 % | 28% |
| Factor VII | 20 - 50 % | 27% |
| Factor X | 20 - 50 % | 30% |
| Factor VII+ X | 20 - 50 % | 24% |
| Factor VIII | 20 - 50 % | 31% |
| Factor IX | 20 - 50 % | 29% |
| Factor XI | 20 - 50 % | 27% |
| Factor XII | 20 - 50 % | 34% |
| Factor XIII | 20 - 50 % | 49% |
| Free Protein S (Antigen) | 20 - 50 % | 34% |
| Plasminogen | 20 - 50 % | 40% |
| Protein C Chromogenic | 20 - 50 % | 33% |
| Protein C (Antigen) | 20 - 50 % | 30% |
| Protein C Clotting | 20 - 50 % | 39% |
| Protein S (Activity) | 20 - 50 % | 31% |
| von Willebrand Factor (Antigen) | 20 - 50 % | 41% |

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-07-19

Quality Control Manager: S. LECOURT