



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN APIXABAN CONTROL LOW
Product Code: 225201
Lot No.: FB06931E
Expiration date: 2024-10-08

Components	Quantity	Lot	Expiration Date
C1	6 vials	FB0693DW1E	2024-10-08
C2	6 vials	FB0693DX1E	2024-10-08

Analytical Data	Specification	Result
Concentration	C1: 15 -35 ng/mL C2: 45 -105 ng/mL	27 ng/mL 82 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-05-18

Quality Control Manager:

S.LECOURT