



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description:** BIOPHEN RIVAROXABAN CALIBRATOR LOW  
**Product Code:** 226001  
**Lot No.:** FB07541B  
**Expiration date:** 2024-10-21

Components	Quantity	Lot	Expiration Date
CAL I	4 vials	FB0754DJ1B	2024-10-21
CAL II	4 vials	FB0754DK1B	2024-10-21
CAL III	4 vials	FB0754DL1B	2024-10-21

Analytical Data	Specification	Result
Concentration	CAL I : < 20 ng/ml CAL II : 40 -60 ng/ml CAL III : 80 -120 ng/ml	0 54 96
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2022-06-07

Quality Control Manager: **S.LECOURT**

