



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description:** BIOPHEN ARIXTRA CALIBRATOR  
**Product Code:** 222501  
**Lot No.:** FB085318  
**Expiration date:** 2024-11-05

Components	Quantity	Lot	Expiration Date
CAL 1	3 vials	FB0853CN18	2024-11-05
CAL 2	3 vials	FB0853CO18	2024-11-05
CAL 3	3 vials	FB0853CP18	2024-11-05
CAL 4	3 vials	FB0853CQ18	2024-11-05

Analytical Data	Specification	Result
Concentration	CAL 1: $\leq 0.05 \mu\text{g/ml}$ CAL 2: $0.40 - 0.60 \mu\text{g/ml}$ CAL 3: $0.90 - 1.10 \mu\text{g/ml}$ CAL 4: $1.35 - 1.60 \mu\text{g/ml}$	0 0.51 1.01 1.48
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2022-06-21

Quality Control Manager: S.LECOURT