



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN RIVAROXABAN CONTROL PLASMA
Product Code: 224501
Lot No.: FB09101C
Expiration date: 2024-11-16

Components	Quantity	Lot	Expiration Date
C1	6 vials	FB0910BD1C	2024-11-16
C2	6 vials	FB0910BE1C	2024-11-16

Analytical Data	Specification	Result
Concentration	C1: 50 -150 ng/mL C2: 250 -350 ng/mL	C1: 100 ng/mL C2: 309 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-08-25

Quality Control Manager:

S.LECOURT