



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : PT-Phen™ LRT
Product Code : CK586K-CK586L
Lot No. : FB1844BX
Expiration date : 2024-11-29

| Components | Quantity | Lot | Expiration Date |
|-------------------|---------------------------------------|------------|-----------------|
| R1 : PT-Phen™ LRT | 6 vials (CK586K) 12 vials (CK586L) | FB1844doBX | 2024-11-29 |

| Analytical Data | Specification | Result |
|---------------------|---------------|-------------------|
| PT on Normal Plasma | 11 ± 1 sec | 10.6 sec (on CS) |
| ISI (WHO Method) | 1.0 ± 0.1 | 1.01 (on CS) |

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-01-10

Quality Control Manager :

S.LECOURT