



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description:** RISTOCETIN  
**Product Code:** AG004K  
**Lot No.:** FB1042  
**Expiration date:** 2024-12-13

Components	Quantity	Lot	Expiration Date
RISTOCETIN	3 vials	FB1042	2024-12-13

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 1.2mg/ml on CS	$\geq 50\%$	91%
Reactivity (% AgregMax) at 1.2mg/ml on Agregometer	$\geq 50\%$	106%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2022-07-06

Quality Control Manager: S. LECOURT

