



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN LMWH CONTROL PLASMA
Product Code : 223001
Lot No. : FB1518
Expiration date : 2025-01-20

Components	Quantity	Lot	Expiration Date
C3	6 vials	FB1263BO2G	2025-01-20
C4	6 vials	FB1262BP2D	2025-01-20

Analytical Data	Specification	Result
Concentration	C3 : 0.70 -0.90 IU/ml C4 : 1.05 -1.35 IU/ml	0.83 IU/mL 1.24 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-09-12

Quality Control Manager :

S.LECOURT

