



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN LMWH CONTROL PLASMA  
**Product Code :** 223001  
**Lot No. :** FB2015  
**Expiration date :** 2025-01-20

Components	Quantity	Lot	Expiration Date
C3	6 vials	FB1812BO2H	2025-04-25
C4	6 vials	FB1262BP2D	2025-01-20

Analytical Data	Specification	Result
Concentration	C3 : 0.70 -0.90 IU/ml C4 : 1.05 -1.35 IU/ml	0.78 IU/mL 1.24 IU/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2022-12-02

Quality Control Manager : S.LECOURT

