



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN UFH CONTROL C2
Product Code: 223901
Lot No.: FB12301C
Expiration date: 2025-01-11

Components	Quantity	Lot	Expiration Date
C2	12 vials	FB1230BY1C	2025-01-11

Analytical Data	Specification	Result
Concentration	C2 : 0.40 -0.60 IU/mL	0.49 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-08-23

Quality Control Manager: S.LECOURT