



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : PT-Phen™ LRT
Product Code : CK586K-CK586L
Lot No. : FC0279DA
Expiration date : 2025-02-22

Components	Quantity	Lot	Expiration Date
R1 : PT-Phen™ LRT	6 vials (CK586K) 12 vials (CK586L)	FC0279doDA	2025-02-22

Analytical Data	Specification	Result
PT on Normal Plasma	11 ± 1 sec	11.2 (on CS) 11.2 (on CN)
ISI (WHO Method)	1.0 ± 0.1	1.03 (on CS) 1.01 (on CN)

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-03-24

Quality Control Manager :

S.LECOURT