



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description : ZYMUTEST Protein C**

**Product Code : RK027A**

**Lot No. : FB1398**

**Expiration date : 2025-02-21**

Components	Quantity	Lot	Expiration Date
Anti (h) Protein C pre-coated plate	12x8 wells	FB1494	2025-03-07
Protein C Standard	3 vials	FB1413	2025-02-21
Protein C Control I	1 vial	FB1415	2025-02-21
Protein C Control II	1 vial	FB1416	2025-02-21
Anti-(h)-Protein C HRP immunoconjugate	3 vials	FB1495	2025-03-10
Sample diluent	2x50 ml	FB1503	2026-03-02
Wash solution	1x50 ml	FB0866	2025-11-03
Conjugate diluent	1x25 ml	FB1456	2026-02-26
TMB	1x25 ml	210604D01	2025-06-30
Sulfuric Acid 0,45M	1x6 ml	FB0506	2025-09-01

Analytical Data	Specification	Result
Reactivity	A450 for 100% $\geq$ 1.40	1.73
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.023
Detection threshold	$\leq$ 5 %	< 1%

### Calibration Curve for information

						Acceptance Range	MV
Conc.	C	C/2	C/4	C/10	C/20	CI [ 60-82 ]	67
OD	1.903	1.366	0.872	0.437	0.261	CII [ 35-52 ]	43

MV = Measured value

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2022-10-18

Quality Control Manager : S.LECOURT

