



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description : BIOPHEN HEPARIN LRT**

**Product Code : 221011**

**Lot No. : FC2018G2**

**Expiration date : 2025-03-25**

Components	Quantity	Lot	Expiration Date
R1 : SXa-11 substrate	4 vials	FC2018BHG2	2025-08-22
R2 : Bovine FXa	4 vials	FC2018BIG2	2025-03-25

Analytical Data	Specification	Result
SUPERIMPOSITION REACTIVITY, LINEARITY, ACCURACY ON STAR	<ul style="list-style-type: none"> <li>For a same A405 (UFH/LMWH):</li> <li>MV = TV <math>\pm</math>0.05 IU/ml for conc. <math>\leq</math> 0.50 IU/ml</li> <li>MV = TV <math>\pm</math>0.10 IU/ml for conc. <math>&gt;</math> 0.50 IU/ml</li> <li><math>R^2 \geq 0.98</math></li> </ul>	PASS PASS >0.99
SUPERIMPOSITION REACTIVITY, LINEARITY, ACCURACY ON CS	<ul style="list-style-type: none"> <li>For a same A405 (UFH/LMWH):</li> <li>MV = TV <math>\pm</math>0.05 IU/ml for conc. <math>\leq</math> 0.50 IU/ml</li> <li>MV = TV <math>\pm</math>0.10 IU/ml for conc. <math>&gt;</math> 0.50 IU/ml</li> <li><math>R^2 \geq 0.98</math></li> </ul>	PASS PASS >0.99
DETECTION THRESHOLD	$\leq 0.05$ IU/ml	<0.01 IU/mL

### Calibration Curve for information

Conc. LMWH	0.00	0.44	0.93	1.39	1.81
OD STAR	1.745	1.207	0.853	0.611	0.437
OD CS	0.754	0.488	0.321	0.214	0.148

	Acceptance Range	MV
C3	[0.68 - 0.88]	0.76
C4	[1.07 - 1.27]	1.17

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

**Quality Control Approval**

Quality Control Manager :

13 NOV. 2023

S. LECOURT

