



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN HEPARIN LRT

Product Code : 221013

Lot No. : FC2017G1

Expiration date : 2025-03-25

Components	Quantity	Lot	Expiration Date
R1 : SXa-11 substrate	3 vials	FC2017BHG1	2025-08-22
R2 : Bovine FXa	3 vials	FC2017BIG1	2025-03-25

Analytical Data	Specification	Result
SUPERIMPOSITION REACTIVITY, LINEARITY, ACCURACY ON STAR	<ul style="list-style-type: none"> For a same A405 (UFH/LMWH): MV = TV \pm0.05 IU/ml for conc. \leq 0.50 IU/ml MV = TV \pm0.10 IU/ml for conc. $>$ 0.50 IU/ml $R^2 \geq 0.98$ 	PASS PASS >0.99
SUPERIMPOSITION REACTIVITY, LINEARITY, ACCURACY ON CS	<ul style="list-style-type: none"> For a same A405 (UFH/LMWH): MV = TV \pm0.05 IU/ml for conc. \leq 0.50 IU/ml MV = TV \pm0.10 IU/ml for conc. $>$ 0.50 IU/ml $R^2 \geq 0.98$ 	PASS PASS >0.99
DETECTION THRESHOLD	\leq 0.05 IU/ml	<0.01 IU/mL

Calibration Curve for information

Conc. LMWH	0.00	0.44	0.93	1.39	1.81
OD STAR	1.745	1.207	0.853	0.611	0.437
OD CS	0.754	0.488	0.321	0.214	0.148

	Acceptance Range	MV
C3	[0.68 - 0.88]	0.76
C4	[1.07 - 1.27]	1.17

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

13 NOV. 2023

Quality Control Manager :

S. LECOURT

