



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN HEPARIN CALIBRATOR
Product Code : 222001
Lot No. : FB15811E
Expiration date : 2025-03-25

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FB1581BJ1E	2025-03-25
CAL 2	4 vials	FB1581BK1E	2025-03-25
CAL 3	4 vials	FB1581BL1E	2025-03-25
CAL 4	4 vials	FB1581BM1E	2025-03-25
CAL 5	4 vials	FB1581BN1E	2025-03-25

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 0.05 IU/ml CAL 2 : 0.30 -0.50 IU/ml CAL 3 : 0.70 -0.90 IU/ml CAL 4 : 1.10 -1.30 IU/ml CAL 5 : 1.45 -1.75 IU/ml	0.00 IU/ml 0.39 IU/ml 0.81 IU/ml 1.17 IU/ml 1.56 IU/ml
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-10-26

Quality Control Manager : **S.LECOURT**

