



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** HEMOCLOT™ LA-S  
**Product Code :** CK090K  
**Lot No. :** FB1544B3  
**Expiration date :** 2025-03-14

Components	Quantity	Lot	Expiration Date
HEMOCLOT LA-S	6 vials	FB1544g5B3	2025-03-14

Analytical Data	Specification	Result
Normal Plasma	Mean CT < 45 sec NR < 1.20	37 sec (CS) 38 sec (STAR) PASS
Pathological Plasma	NR ≥ 1.20	PASS

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2022-10-12

Quality Control Manager :

S.LECOURT