



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: Factor VIII:C Deficient Plasma
Product Code: DP040A/K
Lot No.: FB15621H
Expiration date: 2025-03-15

Components	Quantity	Lot	Expiration Date
Factor VIII:C Deficient Plasma	1 vial (DP040A) 6 vials (DP040K)	FB1562JV1H	2025-03-15

Analytical Data	Specification	Result
Calibration curve	CT (0-100%) \geq 40 sec CT (100%) $<$ 75 sec	76 sec 60 sec
Concentration FVIII	$<$ 1%	$<$ 1%
Concentration FV	$>$ 50 %	56%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-10-06

Quality Control Manager: **S.LECOURT**