




Product name:	Protein L ligand leakage ELISA kit		
Production date:	2023-10		
Date of release:	2023-11-20		
Expiry date:	2025-04		
Store:	+4°C. Protein L reference at -20°C or -80°C.		
Contents:	ELISA kit for detection and quantification of ligand leakage in eluates from Protein L affinity chromatography medium containing immunoglobulin (Ig) or Ig-fragments.		
	Kit contents	Lot	Expiry date
	1 x Anti-Protein L coated microtiter plate	294004	2025-04
	130 µl HRP-conjugated anti-Protein L pAb	294324	2025-04
	1 x 200 µl Protein L reference (600 ng/ml) (10-0027)	285004	2024-12
	5 x 200 µl Protein L reference (600 ng/ml) (10-0028)	285004	2024-12
	13 ml EC-Blue Enhanced™	293801-04	2026-03
	7 ml Stop solution, 0.5 M H ₂ SO ₄	284002	2027-10
	6 x 5 g PBS Tween tablets	283806-03	2025-09

Quality Control

Analysis	Specification	Result
%CV	< 12%	Passed
Absorbance 30 ng/ml	> 2.1	Passed

	OD 450 nm
Calibrator 0.47 ng/ml	0.173
Calibrator 0.94 ng/ml	0.267
Calibrator 1.88 ng/ml	0.496
Calibrator 3.75 ng/ml	0.986
Calibrator 7.5 ng/ml	1.826
Calibrator 15 ng/ml	2.673
Calibrator 30 ng/ml	2.903

	Concentration
Control sample A	0.239
Control sample B	0.833
Control sample C	2.083



The above material has met all quality specifications and has been reviewed by a quality representative

Sebastian Määttä
 Quality Assurance, Sebastian Määttä

2023-11-20
 Date