



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN LMWH CONTROL C3  
**Product Code :** 223801  
**Lot No. :** FB18122H  
**Expiration date :** 2025-04-25

Components	Quantity	Lot	Expiration Date
C3	12 vials	FB1812BO2H	2025-04-25

Analytical Data	Specification	Result
Concentration	C3 : 0.70 -0.90 IU/ml	0.78 IU/mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2022-11-28

Quality Control Manager : S.LECOURT