



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN FXIa CONTROL SET  
**Product Code :** 224801  
**Lot No. :** FB173817  
**Expiration date :** 2025-04-14

Components	Quantity	Lot	Expiration Date
C1	3 vials	FB173817	2025-04-14
C2	3 vials	FB173817	2025-04-14

Analytical Data	Specification	Result
Concentration	C1 : about 10 mIU/ml C2 : about 30 mIU/ml	13 mIU / mL 33 mIU / mL

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : 2022-12-22

Quality Control Manager : S.LECOURT