



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN ANTITHROMBIN 2.5
Product Code : 221102
Lot No. : FB1921C1
Expiration date : 2025-05-16

Components	Quantity	Lot	Expiration Date
R1 : Bovin Xa	2 vials	FB1921FQC1	2025-05-23
R2 : Substrate	2 vials	FB1921FRC1	2025-05-17
R3: Tris heparin Buffer	2 vials	FB1921cyC1	2025-05-16

Analytical Data	Specification	Result
Calibration curve : reactivity, linearity, slope and accuracy	A405 (0%) = 1.50 ± 0.20 Delta A405 (0-100%) ≥ 0.80 $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	1.695 1.142 0.98 PASS
Detection Threshold	$\leq 5\%$	< 1 %

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-01-16

Quality Control Manager : S.LECOURT