



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CONTROL PLASMA
Product Code : 224501
Lot No. : FB18481D
Expiration date : 2025-05-04

Components	Quantity	Lot	Expiration Date
C1	6 vials	FB1848BD1D	2025-05-04
C2	6 vials	FB1848BE1D	2025-05-04

Analytical Data	Specification	Result
Concentration	C1 : 50 -150 ng/ml C2 : 250 -350 ng/ml	99 ng/mL 321 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2022-12-19

Quality Control Manager : S.LECOURT

P/O