



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description :	Factor IX Deficient Plasma
Product Code :	DP050A/K
Lot No. :	FB19541H
Expiration date :	2025-05-27

Components	Quantity	Lot	Expiration Date
Factor IX Deficient Plasma	1 vial (DP050A) 6 vials (DP050K)	FB1954JW1H	2025-05-27

Analytical Data	Specification	Result
Calibration curve	CT (0-100%) ≥ 40 sec CT (100%) < 80 sec	86.0 sec 53.7 sec
Concentration FIX	< 1%	< 1%
Concentration FV	> 50 %	91%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-01-05

Quality Control Manager : **S.LECOURT**

