



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : PT Calibrator Set
Product Code : SC084K
Lot No. : FB191915
Expiration date : 2025-05-19

Components	Quantity	Lot	Expiration Date
CAL 1	2 vials	FB1919di15	2025-05-19
CAL 2	2 vials	FB1919dj15	2025-05-19
CAL 3	2 vials	FB1919dk15	2025-05-19
CAL 4	2 vials	FB1919dl15	2025-05-19
CAL 5	2 vials	FB1919dm15	2025-05-19

Analytical Data	Specification	Result
CAL 1	INR : 0.9-1.1	PASS
CAL 2	INR : 1.8-2.2	PASS
CAL 3	INR : 2.6-3.4	PASS
CAL 4	INR : 3.4-4.6	PASS
CAL 5	INR : 4.3-5.7	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-03-28

Quality Control Manager : **S.LECOURT**

