




A Company of Sysmex Group

 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Item Description : LIAPHEN Fibrinogen

Item Code : 120102

Lot No. : FB19561H

Expiration date : 2025-06-06

Components	Quantity	Lot	Expiration Date
R1 : Latex reagent	4 vials	FB1956JA1H	2025-06-06

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	Delta A620 (0-20µg/ml) \geq 0.40 Controls measured within acceptance ranges $R^2 \geq$ 0.98	0.558 PASS PASS
Limit of Quantification	\leq 1 µg/ml (in the test dilution (1:200))	< 1 µg/mL
Hook effect	> 12 g/L (in the test dilution (1:200))	> 12g/L

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-01-10

Quality Control Manager : **S.LECOURT**

