



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CALIBRATOR

Product Code : 222701

Lot No. : FC00291C

Expiration date : 2025-06-21

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FC0029BA1C	2025-06-21
CAL 2	4 vials	FC0029BB1C	2025-06-21
CAL 3	4 vials	FC0029BC1C	2025-06-21

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/ml CAL 2 : 200 -300 ng/ml CAL 3 : 400 -500 ng/ml	0 ng/mL 245 ng/mL 454 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2023-01-20

Quality Control Manager : **S.LECOURT**

