

Certificate of Analysis

Product: 5-PLASMA Normal Plasma 5

Reference Number: 5D-40401B

Lot Number: 335500

Expiration Date: 2025-06-22

Storage: 2-8°C

Presentation: 8 vials of 5 mL

Specifications:

Analytical Data	Specifications	Result
INR	< 1,3	1.18
aPTT ratio	≤ 1.20	1.04
Fibrinogen	≥ 2 g/L	2.65 g/L
FVIII:C	> 50%	102%
LA DVV test / DVV confirm ratio	≤ 1.20	1.13

For Research Use Only. Not for use in diagnostic procedures. For *in vitro* use only.

Signature:



Urs-Peter Geiger
Date: 2023-08-31