



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

CERTIFICATE OF ANALYSIS	
<b>Product Description :</b>	<b>CEPHEN 2.5</b>
<b>Product Code :</b>	<b>CK512K</b>
<b>Lot No. :</b>	<b>FC1502F4</b>
<b>Expiration date :</b>	<b>2025-06-22</b>

Components	Quantity	Lot	Expiration Date
CEPHEN 2.5	6 vials	FC1502KBF4	2025-06-22

Analytical Data	Specification	Result
aPTT on Normal Plasma	Ratio < 1.20	0.99 on STAR 0.99 on CS 0.99 on CN

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval**      Quality Control Manager :

21 AOUT 2023

S. LECOURT

