



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: ZYMUTEST (h) PAI-1 Antigen
Product Code: RK012A
Lot No.: FB0629
Expiration date: 2025-06-30

Components	Quantity	Lot	Expiration Date
Anti-(h)-PAI-1 pre-coated plate	12x8 wells	FB0786	2025-10-20
(h) PAI-1 standard	3 vials	FB0744	2025-10-15
(h) PAI-1 Control I	1 vial	FB0746	2025-10-15
(h) PAI-1 Control II	1 vial	FB0747	2025-10-15
Anti-(h)-PAI-1 HRP immunoconjugate	3 vials	FB0787	2025-10-29
F-Sample diluent	2x50 ml	FB0631	2025-09-22
Wash solution	1x50 ml	FB0445	2025-08-18
Conjugate diluent	1x25 ml	FB0595	2025-09-15
TMB	1x25 ml	210604D01	2025-06-30
Sulfuric Acid 0,45M	1x6 ml	FB0506	2025-09-01

Analytical Data	Specification	Result
Reactivity	A450 for 8 ng/ml \geq 1.40	1.40
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.002
Detection threshold	\leq 0.5 ng/mL	< 0.2

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	1.618	1.103	0.634	0.276	0.142

	Acceptance Range	MV
CI	[29.6 - 40.0]	34.8
CII	[9.3 - 14.0]	11.7

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: 2022-06-22

Quality Control Manager :

S.LECOURT

