



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

CERTIFICATE OF ANALYSIS	
<b>Product Description:</b>	<b>ZYMUTEST vWF</b>
<b>Product Code:</b>	<b>RK030A</b>
<b>Lot No.:</b>	<b>FB0917</b>
<b>Expiration date:</b>	<b>2025-06-30</b>

Components	Quantity	Lot	Expiration Date
Anti (h) vWF pre-coated plate	12x8 wells	FB1025	2025-12-01
vWF standard	3 vials	FB0964	2025-11-20
vWF Control I	1 vial	FB0965	2025-11-20
vWF Control II	1 vial	FB0966	2025-11-20
Anti-(h)-vWF HRP immunoconjugate	3 vials	FB1044	2025-12-08
Sample diluent	2x50 ml	FB0774	2025-10-20
Wash solution	1x50 ml	FB0445	2025-08-18
Conjugate diluent	1x25 ml	FB0764	2025-10-16
TMB	1x25 ml	210604D01	2025-06-30
Sulfuric Acid 0,45M	1x6 ml	FB0237	2025-07-16

Analytical Data	Specification	Result
Reactivity	A450 for 100 % $\geq 1.40$	1.45
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	$< 0.100$	0.011
Detection threshold	$\leq 5\%$	$< 1\%$

*Calibration Curve for information*

Conc.	C	C/2	C/4	C/10	C/20
OD	1.791	1.258	0.768	0.350	0.191

	Acceptance Range	MV
CI	[ 68 - 92 ]	81
CII	[ 35 - 53 ]	48

MV = Measured value

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval: 2022-07-18

Quality Control Manager: S.LECOURT